



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
 T: +44 (0) 1787 882038 | F: +44 (0) 1787 882039 | E: info@modusgauges.com | W: modusgauges.com

UKAS TRACEABLE CALIBRATION CERTIFICATE NO: 30447

TO : INSTRUMENT SERIAL NO : 13CK425368

OUR JOB NO : 16060

ITEM	QTY	INSTRUMENT DESCRIPTION WITH IDENTIFICATION/REF/PART CODE
1	1	100MM DIAMETER PRESSURE GAUGE, STAINLESS STEEL CASE, STAINLESS STEEL INTERNALS/ WETTED PARTS, DIRECT BOTTOM ENTRY, 3/8 BSP THREAD, 0-11 BAR/PSI, CLASS 1.0, DRY, PRODUCT CODE B462A348L

INSTRUMENT READING SCALED BAR	TEST PRESSURE READING RISING BAR	TEST PRESSURE READING FALLING BAR
0	0	0
2	1.981	1.928
4	3.964	3.905
6	5.986	5.919
8	8.027	7.992
10	10.064	10.041
11	11.066	-

We certify that the above instrument has been tested to BS EN 837-1: 1998. Accuracy tested to and within ± 1.00% FSD limits. Readings are as shown above.

Instrument tested with: WATER

Instrument tested in: VERTICAL position

Laboratory Temperature 20°C ± 2°C

Calibration Equipment Used:

Transmitter:	0-20 Bar
Serial No:	34181
Accuracy:	0.05%

Calibration equipment above is within its period of calibration and traceable to UKAS/NPL/National Standards.

UKAS Lab No: 0173

Signed:

Date: